



EU type-examination certificate
FI 22.1.01 Revision 1



CERTIFICATE

Issued by	Inspecta Tarkastus Oy Notified Body Number 0424	
In accordance with	The Council Directive 2014/31/EU on Non-automatic Weighing Instruments and EN45501:2015	
Issued to	VETEK WEIGHING AB, Hantverkarsvägen 15, S-76493 Väddö, Sweden	
Instrument type	A graduated, self-indicating, electronic, non-automatic weighing instrument, not intended to be used for direct sales to the public	
Type designation	AV series	
Reference No.	230106-1	
Annex	Descriptive annex of type examination certificate (2 pages)	
Remarks	This revision replaces the earlier version.	
Valid until	23 rd of May 2032	
Date of issue	Helsinki, 24 th of May 2023	
Signatories	 Juha Saastamoinen Technical Expert	 Ari Paaajanen Leading Expert

DESCRIPTIVE ANNEX
Of EU type-examination certificate of a non-automatic weighing instrument**CERTIFICATE HISTORY**

CERTIFICATE NO.	DATE	DESCRIPTION
FI 22.1.01	23.5.2022	Type examination certificate first issued
FI 22.1.01 Revision 1	24.5.2023	Format of software version indication updated

NAME AND TYPE OF INSTRUMENT

The weighing instrument is intended for industrial weighing. The type name of instrument is "AV- series"

DESCRIPTION**Construction**

Indicator

Manufacturer	Type	Test certificate number
Vetek Weighing Ab	TI-2000	FIT 01.I.02

Load cell(s):

Any load cell(s) may be used for instruments under this EC type-approval Certificate, provided the following conditions are met:

1. There is a respective OIML Certificate of Conformity (R60) or a test certificate (EN 45501) issued for the load cell by a Notified Body responsible for type examination under Directive 2014/31/EU.
2. The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules (EN 45501:2015 clause F.4), and any particular installation requirements. A load cell marked NH is allowed only if humidity testing to EN 45501 has been conducted on this load cell.
3. The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in EN 45501:2015 clause F.4, at the time of EU verification or declaration of EU conformity of type.
4. The load transmission must conform to one of the examples shown in the WELMEC Guide for load cells.

Devices

Devices are described in the test certificates involved.

TECHNICAL DATA

Accuracy class	III or IIII
Number of scale intervals n	$n \leq 3000$
Software version	10.00xx or 0618xx (x = 0...9)

Software version is indicated during switch on.

More technical data is described in the test certificates involved.

PERIPHERAL DEVICES AND INTERFACES

Peripheral devices and interfaces are described in the test certificates involved

LOCATION OF SEALS AND VERIFICATION MARKS**Sealing**

Sealing is described in the test certificates involved and the load cell cable or the junction box has to be sealed when applicable.

APPROVAL CONDITIONS

No parts of weighing instrument, no matter if they are mentioned or not, may not be in conflict with the essential requirements of the Annex I of the Directive 2014/31/EU.

LOCATION OF CE MARK, SEALS AND INSCRIPTIONS

A sticker with the CE marking of conformity with supplementary metrology marking is placed near the data plate. The supplementary metrology marking shall consist of the capital letter 'M' and the last two digits of the year of its affixing, surrounded by a rectangle. The height of the rectangle shall be equal to the height of the CE marking (5 mm).

The instrument bears markings according to the Annex III in Council Directive 2014/31/EU in such a way that the markings can't be removed. Max, Min and e are marked near the indication device.

The weighing instrument bears the number of EU-type examination certificate, manufacturer mark or name, postal address, accuracy class and serial number.